

**1 Explanation of symbols**

	See chapter <b>2. Safety</b>
	Important information for the user
	Product can be sterilized in a steam sterilizer (Autoclave)
	Can be used in thermal disinfectors
	CE-mark – Confirms that this product fulfills all requirements for medical products
<b>REF</b>	Reference number
<b>SN</b>	Serial number


**2 Safety**


**2.1 Description of danger levels**

	<b>CAUTION</b>
	<b>CAUTION</b> is used if a deficiency in adequate care may result in an endangering of the patient, the user or others.
	<b>WARNING</b>
	<b>WARNING</b> is used if improper use may result in serious injuries.

**2.2 Safety notes**

	<b>WARNING</b>
	<b>Risk of injury for the patient and the operator</b> Concerning damages like irregular running noise, irregular vibrations, an unspecific rise in temperature and/or other defects. → Stop operating immediately and contact our service.
	<b>WARNING</b>
	<b>Risk of infection</b> In case of an injury of the tissue in the oral area, do not proceed to work with air-operated instruments because of a high risk of an infection.

 WARNING	<b>WARNING</b>
	<b>Burns due to an increased heat of the head of the instrument</b> Burns in the oral area may occur due to an increased heat of the head of the instrument.

 CAUTION	<b>CAUTION</b>
	<b>Accelerated deterioration and dysfunctions due to long term non-usage</b> → Before storage, clean, maintain and dry the instrument according to the norm. Store in the original packing, dry and dustproof.

### 3 Product description

This manual can be used for the following Benco products:

1:1 SHEATH [4113-915], 4:1 SHEATH [4104-336]
--

#### 3.1 Intended purpose

This medical product may only be used for the treatment of patients in dental surgeries.

#### 3.2 Use according to regulations

This product may only be used for the intended purposes, executed by professionally trained personnel.

The indications and warnings in this manual must be considered. This also includes the observing of the processing (see chapter **5 Processing according to ISO 17664**) and the valid national health and safety regulations.


Any other use may result in an endangering of the patient, the operator or others and is strictly prohibited.

Before each application, the functionality of the instrument must be tested.

#### 3.3 Technical specifications

	1:1 SHEATH [4113-915]	4:1 SHEATH [4104-336]
Transmission/reduction	1:1	4:1
Max. speed	40 000 min <sup>-1</sup>	40 000 min <sup>-1</sup>
Max. drive speed	40 000 min <sup>-1</sup>	40 000 min <sup>-1</sup>
Label	blue ring	blue ring
Motor connection according to ISO 3964/DIN 13940	yes	yes

#### 3.4 Transport and storage requirements




 CAUTION	<b>CAUTION</b>
	<b>Dysfunction of the instrument while operating after an overcooled storage</b> → Heavily cooled instruments must be warmed to room temperature prior to operating.

For transportation and storage, please consider the following:






Temperature	-20 °C - +70 °C (-4 °F - +158 °F)
Relative humidity	5% - 95%, non-condensing
Air pressure	700 hPa - 1060 hPa (10 psi - 15 psi)

**Protect this medical instrument from moisture!**

## 4 Operation & Usage


 WARNING	<b>WARNING</b>
	<b>Risk of infection for patients and operator due to non-sterile instruments</b> → Before the first and after each further use the instrument must be processed and sterilized as defined in chapter 5 <b>Processing according to ISO 17664.</b>
 WARNING	<b>WARNING</b>
	<b>Risk of infections for others at waste disposal</b> → Before waste disposal, this medical product has to be processed according to instructions and must at least be sterilized.
 CAUTION	<b>CAUTION</b>
	<b>Dysfunctions and/or risk of an infection caused by moist and contaminated compressed air and by contaminated cooling water</b> → This product may only be operated with CE-certified dental units or with others certified by actual national guidelines and with a corresponding supply of water and compressed air according to the valid norms of dental care.

### 4.1 Insertion and removal of the bur

 WARNING	<b>WARNING</b>
	<b>Risk of injury for the patient due to a decoupling of the rotating instrument</b> → Check the holding force of the spindle by pulling at the bur after each insertion of burs.
 WARNING	<b>WARNING</b>
	<b>Risk of injury for the patient and possible damages of the instrument</b> Use of other instruments than those mentioned below. → Only use rotating instruments that fulfill the requirements mentioned above. → Consider the manual for the intended use of the rotating instrument.
 WARNING	<b>WARNING</b>
	<b>Risk of injury due to the use of worn out rotating instruments</b> Rotating instruments can fall out while operating and thus injure the patient. → Do not use any rotating instruments with a worn out shaft.
 WARNING	<b>WARNING</b>
	<b>Risk of injury for the user due to rotating instruments</b> Cuts and resulting infections may occur due to incorrect use. → Always use gloves or a finger protection.
 WARNING	<b>WARNING</b>
	<b>Danger due to a defective spindle</b> In case of a defective spindle, the bur may fall out and cause injuries. → Check the correct attaching of the spindle and the hold of the rotating instrument by slightly pulling at the bur. → Use gloves or finger protection for insertions and removals of the spindle to prevent injuries and infections.



**i** Only use rotating instruments made of steel or carbide metal that fulfill the EN ISO 1797-1 requirements and also the following criteria:


Diameter of the shaft	2.334 to 2.350 mm
Full length of the shaft	max. 22 mm



- ☞ Press the push-button of the back cap and insert the rotating instrument into the spindle with turning movements as far as it will go.
- ☞ Afterwards release the push-button.
- ☞ Check the proper fitting by slightly pulling at the bur.
- ☞ To remove the head let the bur come to a full stop. By pressing the push-button on the back of the head the bur can be released and then be removed.

#### 4.2 Insert the rubber cups or polishing brushes


 WARNING	<b>WARNING</b> <b>Risk of injury for the patient due to a falling out of the rotating instrument</b> → After each insertion of burs check the holding force of the spindle by pulling on the bur.
	<b>WARNING</b> <b>Risk of injury due to the use of worn out rotating instruments</b> Rotating instruments can fall out while operating and thus injure the patient. → Do not use any rotating instruments with a worn out shaft.
 WARNING	<b>WARNING</b> <b>Risk of injury for the user due to rotating instruments</b> Cuts and resulting infections may occur. → Always use gloves or a finger protection.
	<b>WARNING</b> <b>Danger due to a defective spindle</b> In case of a defective spindle, the bur may fall out and cause injuries. → Check the correct attaching of the spindle by slightly pulling at the bur. → Use gloves or finger protection for insertions and removals of the spindle to prevent injuries and infections.




Snap-on and screw-type polishing attachments (calyces and brushes) can be inserted in accordance with DIN EN ISO 13295.

- ☞ Screw or push the insert onto/into the head.
- ☞ Check that the insert is firmly seated.
- ☞ After movement stopped, unscrew or remove the insert.

#### 4.3 Attaching the instrument to the motor


 WARNING	<b>WARNING</b> <b>Dysfunctions due to a decoupling of the instrument while operating</b> An instrument that is not correctly attached to the coupler may loosen while operating. → Prior to each use check if the instrument and the coupler are correctly attached by slightly pulling at the instrument.
	<b>CAUTION</b> <b>Damages due to an insertion or a removal of the instrument, while the motor is rotating</b> Damages to the collector of the motor → Never unplug or plug the instrument to the motor, when the motor is rotating!



- ☞ Place the instrument exactly on the motor and press firmly until the connection clicks in place.
- ☞ Concerning handpieces with light, control the correct position of the glass rod by turning the instrument.
- ☞ Check the proper fitting of the instrument on the motor by slightly pulling at the instrument.
- ☞ For removal, hold the motor in one hand, while pulling the instrument upwards in a turning motion with the other hand.


### 5 Processing according to ISO 17664

#### 5.1 Processing at the location of use


 WARNING	<b>WARNING</b> <b>Risk of infection due to non-sterile instrument.</b> → Always wear gloves.
---	--


- ☞ Process the instrument within one hour after each treatment in order to prevent piling.
- ☞ Immediately remove any debris of cement, blood or composite.
- ☞ Ensure a dry transport of the instrument to the place of processing.
- ☞ Do not put the instrument in any kind of solution.

## 5.2 Cleaning

	<b>WARNING</b>
	<b>Dysfunctions due to cleanings in an ultrasonic bath</b> → Only clean the instrument as described below.

### 5.2.1 Manual pre-cleaning

	<b>CAUTION</b>
	<b>Risk of infection due to multiple uses of dental brushes</b> → To prevent cross infections always use a new dental brush for each instrument.


	<b>CAUTION</b>
	<b>Insufficient cleaning of the surface and the water- and air pipes</b> → In order to ensure a successful cleaning a thorough manual pre-cleaning, even with the small brush, must be done.

Tools:

- Drinkable water 30 °C ± 5 °C (86 °F ± 10 °F)
- Brush or soft tooth brush
- Dental brush
- 50 ml syringe without cannula


- ☞ To pre-clean the instrument place it under running water and carefully brush debris completely off the surface by using one of the tools mentioned above.
- ☞ To pre-clean the spindle, please insert the enclosed dental brush in the spindle opening. Carefully loosen/remove any residue and debris by circular motions with the brush.
- ☞ Repeat the process several times and each time clean the dental brush with water.
- ☞ To pre-clean the water- and air pipes place the syringe at the lower end of the handpiece and rinse the cavities at least 5 times.

### 5.2.2 Cleaning the visible end of the glass rod of dental turbines with light

	<b>CAUTION</b>
	<b>Risk of scratches due to careless treatment</b> Defects of the glass rod result in a lower light quality. → Clean the glass rod with great care.

- ☞ Clean the visible part of the glass rod with warm water and a soft cloth.
- ☞ Dry with air or with a new soft cloth.
- ☞ Pre-check the glass rod before each use. In case of a damage to the glass rod do not operate and contact BENCO repair service immediately.

## 5.3 Disinfection


	<b>CAUTION</b>
	<b>The use of disinfecting baths and/or of chlorinated disinfectants may lead to defects and dysfunctions of the instruments.</b> → Only clean and disinfect the instrument in a thermal disinfectant.

### 5.3.1 Manual disinfection of the surface

Only use disinfectants whose microbiological effectiveness is ensured by the manufacturer (e.g. VAH/DGHM-registration and CE-mark). Concerning the material compatibility, only disinfectants free of aldehyde and based on alcohol may be used.

- ☞ Apply the disinfectant to a cloth and wipe the instrument clean.
- ☞ Consider the reacting time that is set by the manufacturer of the disinfectant.
- ☞ Follow the manual for the disinfectant.

### 5.3.2 Mechanical cleaning and disinfection of the interior and the exterior

 Benco recommends thermal disinfectors that fulfill the EN ISO 15883-1 requirements and are approved by the manufacturer for the straight- and contra angle, as well as for high speed handpieces. The cleaning should be carried out at a minimum of 55°C (131°F) and for at least 5 minutes, the disinfection at a minimum of 90°C (194°F) and for at least 5 minutes (for an A0-value > 3,000). For the cleaning a mild alkaline cleaning agent with a pH of 9 to 11 is recommended, like e.g. Neodisher® MediClean Forte. For the validation of the process see chapter 6.2 Validation of the processing.

- ☞ For the internal cleaning the adaptors specified by the manufacturer must be used.
- ☞ Repeat the process, if visible contaminations still exist after the treatment in a thermal disinfectant.
- ☞ Before use the instrument must be dry and free of residues.
- ☞ To prevent any kind of impairments of the instrument, make sure after every cycle that the instrument is dry on the inside and the outside.
- ☞ Directly afterwards use the Vector Premium Service Oil [4434-533] to lubricate the instrument.

## 5.4 Process of drying

### 5.4.1 Manual drying

☞ Blow on the instrument with compressed air from the inside and the outside until no visible drops are left.

### 5.4.2 Mechanical drying


The drying process normally is part of the cleaning process of your thermal disinfecter. Please follow the instructions for your thermal disinfecter.

☞ In case of moisture remaining on the instrument after the cleaning process, blow with compressed air until completely dry.

## 5.5 Maintenance with the VECTOR Premium Service Oil VL-S [4434-533]

### 5.5.1 Lubrication of the canal for drive air

Lubricate and maintain the instrument after each mechanical cleaning and before each sterilization process.



☞ For maintenance of the cartridge and bearing screw the enclosed lubrication tool onto the lubrication can.

☞ Insert the lubrication tool into the lower end of the instrument and spray into the drive-air canal for approximately 2 seconds.

☞ Plug the instrument onto your treatment unit and let it run for 15 to 20 seconds to expel excess lubricant from the instrument.

## 6 Sterilization



The sterilization bag has to be big enough for the instrument so that the instrument fits without stretching the bag. Concerning quality and application, the bag has to fulfill all the valid norms and has to be suitable for the sterilization process.

☞ Shrink-wrap each instrument individually into a sterilization bag.

### 6.1 Sterilization with a steam sterilizer (Autoclave) according to EN 13060/ISO 17665-1



CAUTION

#### CAUTION

**Damages due to inappropriate maintenance or service may result in dysfunctions and/or deterioration.**

Reduced lifetime of the product

→ Lubricate the instrument with the VECTOR Premium Service Oil [4434-533] before each sterilization.



CAUTION

#### CAUTION

**Contact corrosion caused by moisture**

Damages of the product

→ After sterilization, immediately remove the instrument from the autoclave.



This medical product has a temperature resistance of up to max. 135 °C (275 °F).

Only apply the following sterilization process:

- Autoclave with a triple pre-vacuum (recommended): minimum of 5 minutes at a minimum of 134°C (273°F)

☞ Pay close attention to the manufacturer's instruction.

### 6.2 Validation of the processing



The validation of the cleaning and of the disinfection was carried out with the device Miele® G 7835 in the program Vario TD with the cleanser Neodisher® MediClean Forte.

Adjustment:

- Pre-rinsing for 1 minute

- Cleaning at 55°C (131°F) for 5 minutes with a dosage of 0.5% Neodisher® MediClean Forte

- Neutralization


- Washing down

- Disinfection at 90°C (194°F) for 5 minutes

The sterilization was validated with the device Tuttnauer® EHS 3870 in a fractional pre-vacuum process at 134°C (273°F) for 5 minutes.

Please note, that the reprocessing procedure performed in your dental office must be validated also.

### 6.3 Maintenance & service

 CAUTION	<b>CAUTION</b>
	<b>Inappropriate maintenance or service may result in dysfunctions and/or deterioration.</b> → Always perform proper maintenance and service.

### 7 Storage

- ☞ Store the cleaned and sterilized instrument in a dry, dark and cool place, sheltered from dust and germs.
- ☞ Please pay attention to the date of expiry of the sterilization liquid.

### 8 Tools & spare parts

Contra angle handpiece without light/with light	1:1 SHEATH [4113-915],4:1 SHEATH [4104-336]
Lubrication tool	VL-AU, VL-AE
Premium Service Oil	VL-S
Premium Service Oil for KaVo® QUATTROCARE	VL-AQ



KaVo® is a registered trademark. The manufacturer does not have any economical connection to the company mentioned above.

### 9 Warranties

Warranties for this product are stated in the Terms of Purchase on [Benco.com](http://Benco.com).

Distributed by

***BencoDental***<sup>™</sup>

800.GO.BENCO • [BENCO.COM](http://BENCO.COM)  
295 CenterPoint Blvd, Pittston, PA 18640

Made in Germany